



- Royal Belgian Institute of Natural Sciences
- Operational Directorate for Natural Environment



- JOMOPANS-Final Meeting



JOMOPANS

JOINT MONITORING PROGRAMME FOR AMBIENT NOISE IN THE NORTH SEA

Alain Norro Royal Belgian Institute of Natural Sciences - OD-Nature

Rue Vautier, 29, 1000 Bruxelles, Belgium



Rijkswaterstaat
Ministerie van Infrastructuur en Waterstaat

A permanent underwater sound monitoring station in Belgian North Sea The Westhinder measurement pile experience

By Alain Norro (*)

* Royal Belgian Institute of Natural Sciences
OD-Nature
Rue Vautier, 29, 1000 Bruxelles, Belgium

JOMOPANS - Project (2018-2021)

- for MSFD purposes D11C2 ambient sound, 13 measuring sites, one in Belgium water
- For the first time, a continuous measurement station was installed at Westhinder measuring pile end May 2019



Photo/A.Norro/RBINS



The Station

RTsys EA-SDA-14 / cabled Hydrophones B&K 8104




Photo/A.Norro/RBINS



The Station

RTsys EA-SDA-14 / cabled Hydrophones B&K 8104

- Cables are used to connect the station to the hydrophone
 One is close to the monopile (4 m depth) the other on is on the seafloor away from the pile and by 12 m depth



The Station

RTsys EA-SDA-14 / cabled Hydrophones B&K 8104

- System is Powered in 12 v (9 -14) ~ 6 w from the pile power station
- Transmission by VHF to the shore of a limited number of data every minute. 95 percentile of 1/3 octave level at 63 Hz, 125 Hz and 1kHz, size of both hard drives

Calibration Chart for Hydrophone Type 8104

Serial No.: **3087939**
 Reference Sensitivity at 250 Hz: $\pm 2\%$ at $23.0 \pm 0.1^\circ\text{C}$
 Including integral cable: ± 0.6 dB
 Voltage Sensitivity (Open Circuit Sensitivity): -206.7 dB re $1 \text{ V}/\mu\text{Pa}$ or 46.3 dB re $1 \text{ mV}/\mu\text{Pa}$
 Charge Sensitivity: 0.370 pC/Pa
 Capacitance (including integral cable): 799.3 pF
 Cable Capacitance: 120 pF/m
 Leakage Resistance: 5 G Ω at $23.0 \pm 0.1^\circ\text{C}$

Measuring Uncertainty
 Sensitivity at 250 Hz: ± 0.25 dB
 Frequency Response at 4 kHz to 200 kHz: ± 1 dB

Frequency Response (at ref. pos.): Individual Free Field Frequency Response Curve attached
 Measured in watertank at $23.0 \pm 0.1^\circ\text{C}$

Summarized Specifications (re 250 Hz)

Frequency Response (Tolerance field excluding measurement uncertainty):
 0.1 Hz to 10 kHz: ± 1.5 dB
 0.1 Hz to 80 kHz: ± 4 dB
 0.1 Hz to 120 kHz: $+4$ dB, -12 dB


Horizontal Directivity 100 kHz: (XY - plane) ± 2 dB
 Vertical Directivity 50 kHz: (XZ - plane) ± 2 dB

For further information, please see <http://www.bk.dk> and Product Data Sheet BP 0517

Date: 7. Oct. 2018, 15:18h Operator: RGT



Physical (mm)



Cable: Shielded low noise with two conductors, waterlocked to ML-C-915
 Weight (including 10m cable): 1.6 kg

Environmental

Operating Temperature range:
 Short term: -40°C to 120°C
 Continuous: -40°C to 80°C
 Change of Sensitivity with Temperature:
 Change: 0 to 0.03 dB/ $^\circ\text{C}$
 Voltage: 0 to -0.04 dB/ $^\circ\text{C}$

Change of Sensitivity with Static Pressure:
 0 to -3×10^7 dB/Pa
 0 to -0.05 dB/atm

Temperature Transient Sensitivity: ≤ 70 Pa/ $^\circ\text{C}$
 (ANSI S2.11-1969; measured with Brüel & Kjær Charge Preampifier Type 2636, LUF 5 Hz)


Allowable Total Radiation Dose: 5×10^7 Rad

Maximum Operating Static Pressure: 4×10^6 Pa (40 atm)

Note: All Values are typical at 25°C (77°F), unless measurement uncertainty or tolerance limit is specified. All uncertainty values are specified at 2 σ (i.e. expanded uncertainty using a coverage factor of 2).

For further information see User manual

* Sensitivity Traceable to:
 DRLM, Danish Primary Laboratory of Acoustics
 NIST, National Institute of Standards and Technology, USA
 1 Pascal = 10^5 dyne/cm²



Frequency [kHz]	Sensitivity [dB re 1 V/μPa]	Frequency [kHz]	Sensitivity [dB re 1 V/μPa]
4.0	-207.1	59.0	-206.0
5.0	-207.5	56.1	-205.4
6.3	-207.7	63.0	-205.5
8.1	-207.3	71.0	-205.9
10.0	-207.0	80.0	-206.6
12.5	-206.7	90.0	-208.4
16.0	-207.6	100.0	-210.9
20.0	-207.6	112.0	-211.3
25.0	-207.8	125.1	-216.0
28.0	-208.3	140.0	-219.3
31.5	-208.1	160.0	-222.9
35.5	-208.0	180.0	-226.6
40.1	-207.0	200.1	-229.6
45.1	-206.0		

Calibration Chart for Hydrophone Type 8104

Serial No.: 3087939

02 2009-12


JOMOPANS-Final Meeting

Calibration




- Hydrophones are fully calibrated after manufacture by B&K
- Acquisition board is calibrated by RTsys after manufacture of the recorder (cross talk & frequency response of every channel)
- The complete system (hydro - amplifier - cable & acquisition board) calibration is further verified at 250Hz with a B&K pistonphone & B&K reference microphone

deployment



Information sent to the coast



```

MP7.BP3_001_MP7JMPBP3001.txt
Westhinder - Measuring pile (UTC time) Channel B - 63 Hz percentile (Hz)
2019-06-01T00:00:00+00:00 125.00
2019-06-01T00:01:00+00:00 125.00
2019-06-01T00:02:00+00:00 125.00
2019-06-01T00:03:00+00:00 125.00
2019-06-01T00:04:00+00:00 125.00
2019-06-01T00:05:00+00:00 125.00
2019-06-01T00:06:00+00:00 124.00
2019-06-01T00:07:00+00:00 124.00
2019-06-01T00:08:00+00:00 124.00
2019-06-01T00:09:00+00:00 124.00
2019-06-01T00:10:00+00:00 124.00
2019-06-01T00:11:00+00:00 123.00
2019-06-01T00:12:00+00:00 123.00
2019-06-01T00:13:00+00:00 123.00
2019-06-01T00:14:00+00:00 123.00
2019-06-01T00:15:00+00:00 123.00
2019-06-01T00:16:00+00:00 123.00
2019-06-01T00:17:00+00:00 123.00
2019-06-01T00:18:00+00:00 123.00

MP7.BP2_001_MP7JMPBP2001.txt
Westhinder - Measuring pile (UTC time) Channel B - 125 Hz percentile (Hz)
2019-06-01T00:00:00+00:00 122.00
2019-06-01T00:01:00+00:00 123.00
2019-06-01T00:02:00+00:00 123.00
2019-06-01T00:03:00+00:00 123.00
2019-06-01T00:04:00+00:00 123.00
2019-06-01T00:05:00+00:00 124.00
2019-06-01T00:06:00+00:00 124.00
2019-06-01T00:07:00+00:00 124.00
2019-06-01T00:08:00+00:00 124.00
2019-06-01T00:09:00+00:00 124.00
2019-06-01T00:10:00+00:00 124.00
2019-06-01T00:11:00+00:00 124.00
2019-06-01T00:12:00+00:00 124.00
2019-06-01T00:13:00+00:00 124.00
2019-06-01T00:14:00+00:00 124.00
2019-06-01T00:15:00+00:00 124.00
2019-06-01T00:16:00+00:00 124.00
2019-06-01T00:17:00+00:00 124.00
2019-06-01T00:18:00+00:00 124.00
2019-06-01T00:19:00+00:00 124.00
2019-06-01T00:20:00+00:00 124.00
2019-06-01T00:21:00+00:00 124.00

MP7.BP1_001_MP7JMPBP1001.txt
Westhinder - Measuring pile (UTC time) Channel B - 1 kHz percentile (Hz)
2019-06-01T00:00:00+00:00 116.00
2019-06-01T00:01:00+00:00 116.00
2019-06-01T00:02:00+00:00 116.00
2019-06-01T00:03:00+00:00 116.00
2019-06-01T00:04:00+00:00 116.00
2019-06-01T00:05:00+00:00 116.00
2019-06-01T00:06:00+00:00 116.00
2019-06-01T00:07:00+00:00 116.00
2019-06-01T00:08:00+00:00 116.00
2019-06-01T00:09:00+00:00 116.00
2019-06-01T00:10:00+00:00 116.00
2019-06-01T00:11:00+00:00 116.00
2019-06-01T00:12:00+00:00 116.00
2019-06-01T00:13:00+00:00 116.00
2019-06-01T00:14:00+00:00 116.00
2019-06-01T00:15:00+00:00 116.00
2019-06-01T00:16:00+00:00 116.00
    
```

- 95 Percentile of 1/3 Octave band 63 Hz & 125 Hz (D11C2) and 1 kHz
- Size of both hard drive
- Raw data logged on 2 * 4 TB HDD on compressed format

Difficulties-lessons learned

- Hydrophones have been trawled away during covid closure including damages to cables
- Cables are re-routed by the inside of the monopile and are better protected against wave action in an remote off-shore location
- Having a back-up recorder is essential during the development/setup phase of the station
- Access to the hard drive/station is difficult due to weather



Acknowledgements



- MDK – Coastal Division-Flemish Hydrography Johan Vercruysse, Broes Benoot
- Officers and Crew of R/V Belgica and R/V Simon Stevin
- Scientific diving team of RBINS and VLIZ
- Bob Rumes for some of the photos

